

Domain	Range	Instrument & methods	Uncertainty
Time	-1,000 ~ 1000 Seconds	SR 620 Direct Time Interval Measurement	2 ns (UTC(TL)) 200 ns (UTC(BIPM))
Frequency	10 MHz ~ 300 MHz	SR 620 Direct Frequency Measurement	2.3 E-11
	1, 5, 10 MHz	FMAS Phase Comparison	2.3 E-12
	1, 5, 10 MHz	SR-620 Phase Comparison	2.3 E-12
	5, 10 MHz	Q-A7 Phase Comparison	2.3 E-13